Application News

No. G305B

Gas Chromatography

Performance Assessment of Nexis™ SCD-2030 Using Sulfur Compounds Recommended by ASTM D5623

ASTM D5623 provides guidelines for the analysis of sulfurcontaining compounds in light petroleum liquids by gas chromatography with a sulfur chemiluminescence detector (SCD). ASTM D5623 applies to petroleum products with a final boiling point of less than 230 °C at standard atmospheric pressure, such as gasoline.

The SCD provides high selectivity and linearity for sulfur-containing compounds and minimizes the quenching effects of hydrocarbons during sample analysis of complex hydrocarbon matrices.

This document introduces the results of using Nexis SCD-2030 to analyze sulfur compounds standards recommended by ASTM D5623.

Y. Nagao, E. Kobayashi



Fig. 1 Nexis™ SCD-2030

Analytical Conditions

Table 1 shows instrument configuration details and analytical conditions used for all evaluations.

In order to prevent adsorption of sulfur compounds at low concentrations, a sample vaporization injector SPL-2030 (P/N: S221-77100-61/-64) was subjected to a deactivation treatment and installed on the system.

LabSolutions GC was used for instrument control and data collection.

Table 1 Instrument Configuration and Analytical Conditions

Main Unit Nexis GC-2030 /AOC-20i plus SH-I-1MS (30 m \times 0.32 mm I.D. df= 4 μ m) *1 Column Detector SCD-2030 Injection Volume 1 μL Injection Mode Split Split Ratio 1:9 Injection Unit Temp 275 °C Carrier Gas Carrier Gas Control Constant Column Flow Mode (2.8 mL/min) Column Temp. 40 °C (3 min) - 10 °C/min – 250 °C (16 min) Interface Temp. 200°C

Furnace Temp. : 850 °C

Detector Gas : H₂ 100 mL/min, N₂ 10 mL/min,

O₂ 12 mL/min, O₃ 25 mL/min

Analysis of Sulfur Calibration Standard

Nineteen discrete sulfur compounds recommended by ASTM D5623 were separated into two groups for optimal chromatogram separation. Ten out of the nineteen compounds were diluted in n-hexane to yield standard 1 (STD1) and the remaining nine sulfur compounds were diluted in toluene to yield standard 2 (STD2). Diphenyl sulfide was added to each STD at a concentration of 1 ppm(w/w) as an internal standard. Chromatograms of the STD1 and STD2 are shown in Fig. 2.




Fig. 2 Chromatograms of 19 Sulfur Compounds (10 ppm)


^{*1} P/N: 227-36011-01

Repeatability and Linearity Evaluation

Nineteen discrete sulfur compounds were diluted to the following concentrations 0.1 ppm, 1 ppm, 10 ppm, and 100 ppm (w/w) according to ASTM D5623. Table 2 illustrates repeatability (%RSD, n=5) and coefficient of determination (R2) for the Nexis SCD-2030 analysis of the sulfur compound standards. The area ratio value of the target compound to the internal standard was used in this evaluation. The results were highly reproducible with excellent linear response for all compounds. Fig. 3 illustrates the calibration curves for 2-propanethiol and methyl disulfide which are representative of the analytes.

Table 2 Repeatability and Linearity of Sulfur Compounds

| No. | Analytes | 0.1 ppm | 1 ppm | 10 ppm | 100 ppm | STD No | R ² (0.1-100 ppm) |
|-----|-------------------------|------------|----------|-----------|------------|-----------|------------------------------|
| 1 | Methyl mercaptan | 7.1 % | 1.8 % | 5.1 % | 3.0 % | 1 | 0.9999 |
| 2 | Ethyl mercaptan | 6.5 % | 1.5 % | 3.7 % | 2.3 % | 1 | 0.9999 |
| 3 | Dimethyl sulfide | 6.8 % | 1.3 % | 3.7 % | 2.4 % | 1 | 0.9999 |
| 4 | Carbon disulfide | 7.3 % | 1.9 % | 4.4 % | 2.7 % | 1 | 0.9999 |
| 5 | 2-Propanethiol | 6.7 % | 1.4 % | 2.9 % | 2.0 % | 1 | 0.9999 |
| 6 | 2-Methyl-2-propanethiol | 9.7 % | 2.8 % | 2.5 % | 1.9 % | 1 | 0.9999 |
| 7 | 1-Propanethiol | 6.3 % | 1.4 % | 2.9 % | 1.9 % | 1 | 0.9999 |
| 8 | Ethyl methyl sulfide | 7.7 % | 1.8 % | 2.8 % | 2.2 % | 1 | 0.9999 |
| 9 | Thiophene | 6.7 % | 1.7 % | 2.8 % | 2.0 % | 1 | 0.9999 |
| 10 | 2-Butanethiol | 8.5 % | 3.4 % | 1.7 % | 2.4 % | 2 | 0.9999 |
| 11 | 2-Methyl-1-propanethiol | 8.5 % | 3.1 % | 1.7 % | 2.3 % | 2 | 0.9999 |
| 12 | Diethyl sulfide | 4.8 % | 2.9 % | 1.9 % | 2.3 % | 2 | 0.9999 |
| 13 | 1-Butanethiol | 9.1 % | 3.8 % | 1.8 % | 2.4 % | 2 | 0.9999 |
| 14 | Methyl disulfide | 1.9 % | 2.6 % | 2.1 % | 2.3 % | 2 | 1.0000 |
| 15 | 2-Methylthiophene | 9.7 % | 3.2 % | 1.8 % | 2.3 % | 2 | 0.9999 |
| 16 | 3-Methylthiophene | 6.0 % | 2.8 % | 1.5 % | 2.4 % | 2 | 0.9999 |
| 17 | Diethyl disulfide | 3.7 % | 2.8 % | 1.8 % | 2.4 % | 2 | 0.9999 |
| 18 | 5-Methylbenzothiophene | 9.4 % | 1.5 % | 1.9 % | 1.8 % | 1 | 0.9999 |
| 19 | 3-Methylbenzothiophene | 9.6 % | 4.1 % | 1.7 % | 2.5 % | 2 | 0.9999 |
| | | | | | | | |






Fig. 3 Calibration Curve for 2-Propanethiol and Methyl Disulfide

■ Long-Term Stability Evaluation

Long-term stability of a detector is one key measure of performance that directly impacts data reliability. The Nexis SCD-2030 is configured with an industry-first horizontal redox cell, which promotes the complete oxidation-reduction reaction of samples by ensuring sufficient reaction area and reaction time in the cell to achieve stable analysis.

A 10 ppm (w/w sulfur) standard 3 (STD3), containing all nineteen sulfur compounds in Table2, was analyzed for sixteen days continuously. Diphenyl sulfide was added to the sample at a concentration of 1 ppm as an internal standard.

Fig. 4 illustrates the results of relative response factor stability and absolute area stability for the four compounds (1-butanethiol, methyl disulfide, diethyl disulfide, 3-methylbenzothiophene) representing boiling points raining from 98 °C to 273 °C. Each plot represents the daily mean value of the analysis result, and the error bar represents the value of three standard deviations (3σ).

The relative standard deviations of the relative response factors and absolute area for 1-butanethiol, methyl disulfide, diethyl disulfide and methylbenzothiophene are shown in Table 3.

The results of this evaluation showed excellent longterm stability and diurnal variation. The Nexis SCD-2030 not only provides a stable relative response but also a stable absolute response over time.

Table 3 Long-Term Stability for Representative Analytes

| Analytes | 1-Butanethiol | Methyl disulfide | Diethyl disulfide | 3-Methyl- benzothiophene |
|------------------------|---------------|---------------------|----------------------|-----------------------------|
| Response Factor RSD(%) | 1.4 | 1.2 | 1.9 | 1.6 |
| Absolute Area RSD(%) | 3.5 | 3.6 | 3.8 | 3.4 |




Fig. 4 Relative Response Factors and **Absolute Peak Area Over 16 Days**

Nexis is a trademark of Shimadzu Corporation or its affiliated companies in Japan and/or other countries.



Shimadzu Corporation www.shimadzu.com/an/

For Research Use Only. Not for use in diagnostic procedures.
This publication may contain references to products that are not available in your country.

First Edition: Mar. 2019 Revision A: Jun. 2019

Please contact us to check the availability of these products in your country.

The content of this publication shall not be reproduced, altered or sold for any commercial purpose without the written approval of Shimadzu See http://www.shimadzu.com/about/trademarks/index.html for details.

Third party trademarks and trade names may be used in this publication to refer to either the entities or their products/services, whether or not they are used with trademark symbol "TM" or "®".

Shimadzu disclaims any proprietary interest in trademarks and trade names other than its own.

The information contained herein is provided to you "as is" without warranty of any kind including without limitation warranties as to its accuracy or completeness. Shimadzu does not assume any responsibility or liability for any damage, whether direct or indirect, relating to the use of this publication. This publication is based upon the information available to Shimadzu on or before the date of publication, and subject to change without notice.